



COMMONWEALTH of VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE


CENTRAL LABORATORY
A Nationally Accredited Laboratory

700 NORTH 5TH STREET
RICHMOND, VIRGINIA 23219
(804) 786-6971

CERTIFICATE OF INSTRUMENT ACCURACY

INSTRUMENT: Intox EC/IR II
SERIAL NUMBER: 010582
CERTIFICATION DATE: September 3, 2013
DATE RECERTIFICATION DUE: March 2, 2014


The above listed instrument was found to meet all requirements for accuracy and performance established by the Department of Forensic Science. This certification is valid from the above listed certification date until the above listed date for recertification.


Melissa Kennedy


Section Supervisor, Breath Alcohol

CITY OF RICHMOND
COMMONWEALTH OF VIRGINIA

The foregoing was acknowledged before me this 13th day of
September, 2013 by Melissa Kennedy


Ciera N. Boxley, Notary Public

My commission expires: August 31, 2014



DBH
NMJ

**Department of Forensic Science
Breath Alcohol Section
Instrument Maintenance History
1/1/1990 To 9/6/2013**

Instrument Serial Number: 010582

As of 06-Sep-13

Maintenance	Certification	Technician	
Date	Date	Initials	Remarks
03-Sep-13	03-Sep-13	DBH (21250)	FIRMWARE/SOFTWARE UPGRADE. CALIBRATED AND CERTIFIED.
18-Mar-13	18-Mar-13	DBH (21250)	CERTIFIED, REPLACED DRY GAS STANDARD.
26-Sep-12	26-Sep-12	DBH (21250)	CERTIFIED. REPLACED DRY GAS STANDARD.
11-Jul-12	11-Jul-12	DBH (21250)	CERTIFIED.
14-Jun-12	14-Jun-12	DBH (21250)	CALIBRATED BAROMETRIC PRESSURE SENSOR, CERTIFIED
11-Jan-12	11-Jan-12	DBH (21250)	CERTIFIED, REPLACED DRY GAS STANDARD
21-Jul-11	21-Jul-11	DBH (21250)	REPLACED DRY GAS STANDARD, CERTIFIED
25-Jan-11	25-Jan-11	RJP (22428)	CERTIFIED.
11-Aug-10	11-Aug-10	RJP (22428)	CERTIFIED.
19-Feb-10	19-Feb-10	RJP (22428)	REPLACED DRY GAS STANDARD. CERTIFIED
26-Aug-09	26-Aug-09	RJP (22428)	CERTIFIED.
23-Apr-09	23-Apr-09	NMJ (22392)	CERTIFIED.
01-Apr-09	01-Apr-09	DBH (21250)	FIRMWARE/SOFTWARE UPGRADE, CALIBRATED AND CERTIFIED.
04-Feb-09	04-Feb-09	NMJ (22392)	CALIBRATED, CERTIFIED AND ACCEPTED.

DBH
PWH

INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010582 Worksheet Start Date 9/3/2013

Location Goochland CO SO

Address 2938 River Rd West Build G Goochland VA 23063

DFS Technician Donald Hall License No. 21250

☐ Laboratory ☒ On-Site

Site Specifications

	Yes	No
1 Secured Environment: Location must not be accessible to the general public.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
2 Environment free from contaminants	<input checked="" type="checkbox"/>	<input type="checkbox"/>
3 Dedicated telephone line: The phone line should be used solely by the instrument.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
4 Adequate Ventilation: Enough space must be provided for the instrument and the printer to allow for adequate ventilation.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
5 Stable temperature controls: The instrument should not be exposed to direct sources of temperature variation.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
6 115VAC power source: Whenever possible instrument power circuit should not be shared with refrigerators, copy machines or other high amperage devices.	<input checked="" type="checkbox"/>	<input type="checkbox"/>
7 Room Condition: The site should be free of excessive moisture, heat and dust.	<input checked="" type="checkbox"/>	<input type="checkbox"/>

Old serial no.	New serial no.	Supplies
E-Seek		Mouthpieces
Keyboard		Certificates of Analysis
Printer		Operator Worksheet
UPS		Other:

Maintenance (if any)
Firmware/software upgrade.

Comments (if any)

Instrument Serial Number 010582

Certification Date 9/3/2013

Instrument Barometer (mm HG) 751

Reference Barometer (mm HG) 753

Reference Barometer(RB)Serial # 009110

RB Calibration Due 5/30/2014

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.245
0.250	0.247	0.239	0.254	Sample 2	0.245
Precision		sample min	sample max	Sample 3	0.244
0.001		0.244	0.245		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.148
0.150	0.148	0.143	0.152	Sample 2	0.148
Precision		sample min	sample max	Sample 3	0.148
0		0.148	0.148		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.079
0.080	0.079	0.076	0.082	Sample 2	0.079
Precision		sample min	sample max	Sample 3	0.079
0		0.079	0.079		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.019
0.020	0.020	0.017	0.023	Sample 2	0.019
Precision		sample min	sample max	Sample 3	0.019
0		0.019	0.019		

all measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

☒ Calibrated☒ Certified☐ Replaced dry gas standard (+O-ring)

Dry gas standard Lot No. (with tank no.)

AG226201-19

☒ Measurement Assurance Check☒ Run a Breath Test (Instrument Test)

(successful completion denotes satisfactory condition of the item)

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.099
0.100	0.099	0.096	0.102	Sample 2	0.099
Precision		sample min	sample max	Sample 3	0.099
0		0.099	0.099		

DFS Technician

Donald B. Hall

Date

9/6/13

Issuing Analyst

Melissa Kennedy

Date

9/10/13

Intox EC/IR II QA Worksheet

Issued by Calibration and Training Program Manager

Issue Date: 4-March-2013

DFS Document 250-F113

Revision 6

Page 2 of 2

DBH
NSK




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 09/03/2013
TEST EQUIPMENT NUMBER 010582		

RESULTS: TIME SAMPLE TAKEN **11:49 EDT**

SAMPLE'S ALCOHOL CONTENT **0.00** GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 03 Sep 2013 at 11:56

TLN
DBH

ink

Test Results

Instrument Serial Number 010582

Test # 000397 Subject Test

Test Location 1 Department of
Test Date 03 Sep 2013

Test Location 2 Forensic Science
Test Time 11:42
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 03 Sep 2013 End Time 11:50

Result Time 11:49

Result Date 03 Sep 2013 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 11:42

Data Type BLK

Sample Value 0.000

Sample Time 11:43

Data Type CHK

Sample Value 0.101

Sample Time 11:44

Data Type BLK

Sample Value 0.000

Sample Time 11:45

Data Type SUBJ

Sample Value 0.000

Sample Time 11:46

Data Type BLK

Sample Value 0.000

Sample Time 11:47

Data Type SUBJ

Sample Value 0.000

Sample Time 11:49

Data Type BLK

Sample Value 0.000

Sample Time 11:49

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG226201-19

Standard Expiration Date 09/18/2014

Tank Pressure 1017

Barometric Pressure 751 mmHg

Blow Sample Number 1 Blow Duration 3.13 sec

Blow Volume 1712 cc End-of-Blow Time 11:46

Blow Sample Number 2 Blow Duration 3.64 sec

Blow Volume 1830 cc End-of-Blow Time 11:49

Tamper Evident Stamp b9d63d01

Test Status Code 0

Test Status Success

Intox EC/IR-II: Software Version

Department of Forensic Science

Serial Number: 010582

System Date: 09/03/2013

System Time: 12:13 EDT

Software Version: VS00333-K

Software CRC: 56745BCA

Firmware Version: VS00171-CL

Firmware CRC: 7107756B

DB/

Net

Intox EC/IR-II: Calibration

Department of Forensic Science

Serial Number: 010582 Test Number: 398

Test Date: 09/03/2013 Test Time: 12:16 EDT

Dry Gas Target: 0.099

Lot Number: 0651210013-05 Exp Date: 04/01/2014

Tank Pressure: 463 psi Barometric Pressure: 750 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	12:16
CAL	0.099	12:17

Test Status: *Success*

Calibration CRC: AADAEA5F

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010582 Test Number: 399

Test Date: 09/03/2013 Test Time: 12:20 EDT

Technician Initials: DBH

Dry Gas Target: 0.247

Lot Number: 29612250I2-10 Exp Date: 11/01/2014

Tank Pressure: 384 psi Barometric Pressure: 750 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	12:20
BLK	0.000	12:21
CHK	0.245	12:22
BLK	0.000	12:23
CHK	0.245	12:24
BLK	0.000	12:26
CHK	0.244	12:27
BLK	0.000	12:28

Success

Standard Value (at sea level): 0.250 g/210L

Certification Status: Continued

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010582 Test Number: 400

Test Date: 09/03/2013 Test Time: 12:31 EDT

Technician Initials: DBH

Dry Gas Target: 0.148

Lot Number: 0651215014-07 Exp Date: 04/01/2014

Tank Pressure: 154 psi Barometric Pressure: 750 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	12:31
BLK	0.000	12:32
CHK	0.148	12:33
BLK	0.000	12:34
CHK	0.148	12:35
BLK	0.000	12:37
CHK	0.148	12:37
BLK	0.000	12:39

Success

Standard Value (at sea level): 0.150 g/210L

Certification Status: Continued

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010582 Test Number: 401

Test Date: 09/03/2013 Test Time: 12:41 EDT

Technician Initials: DBH

Dry Gas Target: 0.079

Lot Number: 0651208012-03 Exp Date: 04/01/2014

Tank Pressure: 345 psi Barometric Pressure: 750 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	12:41
BLK	0.000	12:42
CHK	0.079	12:43
BLK	0.000	12:44
CHK	0.079	12:45
BLK	0.000	12:46
CHK	0.079	12:47
BLK	0.000	12:48

Success

Standard Value (at sea level): 0.080 g/210L

Certification Status: Continued

Intox EC/IR-II: Certification

Department of Forensic Science

Serial Number: 010582 Test Number: 402

Test Date: 09/03/2013 Test Time: 12:50 EDT

Technician Initials: DBH

Dry Gas Target: 0.020

Lot Number: 0651202011-05 Exp Date: 04/01/2014

Tank Pressure: 232 psi Barometric Pressure: 750 mmHg

System Check: Passed

Test	g/210L	Time
DIAG	Pass	12:51
BLK	0.000	12:52
CHK	0.019	12:52
BLK	0.000	12:54
CHK	0.019	12:54
BLK	0.000	12:56
CHK	0.019	12:56
BLK	0.000	12:58

Success

Standard Value (at sea level): 0.020 g/210L

Certification Status: Success

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010582 Test Number: 403

Test Date: 09/03/2013 Test Time: 13:01 EDT

Dry Gas Target: 0.296

Lot Number: AG322702-04 Exp Date: 08/15/2015

Tank Pressure: 522 psi Barometric Pressure: 750 mmHg

System Check: Passed

Test	g/210L	Time
BLK	0.000	13:02
CHK	0.299	13:03
BLK	0.000	13:05
CHK	0.298	13:06
BLK	0.000	13:07
CHK	0.298	13:08

Test Status: Success

Calibration CRC: AADAEA5F

DBH

NSK

Intox EC/IR-II: Accuracy Check

Department of Forensic Science

Serial Number: 010582 Test Number: 404

Test Date: 09/03/2013 Test Time: 13:13 EDT

Dry Gas Target: 0.099

Lot Number: AG226201-19 Exp Date: 09/18/2014

Tank Pressure: 1002 psi Barometric Pressure: 750 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	13:14
CHK	0.099	13:15
BLK	0.000	13:16
CHK	0.099	13:17
BLK	0.000	13:19
CHK	0.099	13:19

Test Status: *Success*

Calibration CRC: AADAEA5F




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY HALL, DONALD, B		AGENCY DFS Central Lab
DFS LICENSE NUMBER 21250	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 09/03/2013
TEST EQUIPMENT NUMBER 010582		

RESULTS: TIME SAMPLE TAKEN **13:28 EDT**

SAMPLE'S ALCOHOL CONTENT **0.00** GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20_____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____
OPERATOR'S SIGNATURE



IntoxNet MIS Report

Report Generated 06 Sep 2013 at 09:37

DBH
M2K

Test Results

Instrument Serial Number 010582

Test # 000405

Subject Test

Test Location 1 Department of
Test Date 03 Sep 2013

Test Location 2 Forensic Science
Test Time 13:21
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name HALL
Agency DFS Central Lab

Operator's First Name DONALD

Operator's Middle Initial B

Card Serial Number 121250

License Number 21250

Effective Date 10/01/2012

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 03 Sep 2013 End Time 13:29

Result Time 13:28

Result Date 03 Sep 2013 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 13:22

Data Type BLK

Sample Value 0.000

Sample Time 13:22

Data Type CHK

Sample Value 0.099

Sample Time 13:23

Data Type BLK

Sample Value 0.000

Sample Time 13:25

Data Type SUBJ

Sample Value 0.000

Sample Time 13:26

Data Type BLK

Sample Value 0.000

Sample Time 13:27

Data Type SUBJ

Sample Value 0.000

Sample Time 13:28

Data Type BLK

Sample Value 0.000

Sample Time 13:29

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG226201-19

Standard Expiration Date 09/18/2014

Tank Pressure 992

Barometric Pressure 750 mmHg

Blow Sample Number 1 Blow Duration 3.47 sec

Blow Volume 1775 cc End-of-Blow Time 13:26

Blow Sample Number 2 Blow Duration 3.69 sec

Blow Volume 1788 cc End-of-Blow Time 13:28

Tamper Evident Stamp 09e757e9

Test Status Code 0

Test Status Success